

**WADSWORTH CENTER  
NEW YORK STATE DEPARTMENT OF HEALTH  
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM**

**Proficiency Test Report**

Lab Id: 11642      USGS NATIONAL WATER LAB  
 POB 25046-MS 407-BLDG 95  
 EPA Lab Code:    DENVER, CO-80225-0046  
 CO00009          United States  
                       (303) 236-2000  
                       Director: MR. GREGORY MOHRMAN

Shipment Date : 27-Jan-2003  
 Closing Date   : 13-Mar-2003  
 Score Date     : 31-Mar-2003

This report may contain data that are not covered by the NVLAP accreditation.  
 \*\* Indicates NVLAP accredited analyte. Lab Code 200387-0. ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 260      Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/Target	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water		Organophosphate Pesticides						
Azinphos methyl EPA Code: N/A	ug/L	6012	Method Not Specified	14.7	12.2		3.06 - 21.3	Satisfactory <i>9 passed out of 11 reported results.</i>
Diazinon EPA Code: N/A	ug/L	6012	Method Not Specified	19	16.9		7.87 - 25.9	Satisfactory <i>11 passed out of 11 reported results.</i>
Disulfoton EPA Code: N/A	ug/L	6012	Method Not Specified	15.9	13.7		6.11 - 21.3	Satisfactory <i>11 passed out of 11 reported results.</i>
Malathion EPA Code: N/A	ug/L	6012	Method Not Specified	6.03	5.15		1.23 - 9.07	Satisfactory <i>13 passed out of 13 reported results.</i>
Parathion ethyl EPA Code: N/A	ug/L	6012	Method Not Specified	13.2	11.6		7.04 - 16.2	Satisfactory <i>13 passed out of 13 reported results.</i>
Parathion methyl EPA Code: N/A	ug/L	6012	Method Not Specified	5	4.93		1.96 - 7.9	Satisfactory <i>13 passed out of 13 reported results.</i>
Sample: Non Potable Water		Chlorinated Hydrocarbon Pesticides						
4,4-DDE ** EPA Code: N/A	ug/L	6017	Method Not Specified	0.316	0.451	0.301 - 0.601	0.226 - 0.676	Satisfactory <i>77 passed out of 97 reported results.</i>