WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Proficiency Test Report

Lab Id: 11642
EPA Lab Code: CO00009

USGS NATIONAL WATER QUALITY LABORATORY
POB 25046-MS 407-BLDG 95
DENVER, CO 80225-0046
(303) 236-2000
Director: MR. MERLE SHOCKEY

Shipment Date : 24-Jan-2006
Closing Date : 09-Mar-2006
Score Date : 29-Mar-2006

This report may contain data that are not covered by the NVLAP accreditation.
** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
NVLAP Lab Code 200387-0.

Shipment: 290 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/ Target *	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water Nitrosamines								
N-Nitrosodimethylamine EPA Code: N/A	ug/L	9014	USGS O-3116-87	144	93.3		9.33 - 191 *62 passed out of 66 reported results.*	Satisfactory
N-Nitrosodi-n-propylamine EPA Code: N/A	ug/L	9014	USGS O-3116-87	58.1	51.2		16.9 - 85.5 *69 passed out of 71 reported results.*	Satisfactory
Sample: Non Potable Water Acid extractables								
2,4,6-Trichlorophenol EPA Code: N/A	ug/L	9035	USGS O-3116-87	138	142		58.2 - 225 *76 passed out of 76 reported results.*	Satisfactory
2,4-Dichlorophenol EPA Code: N/A	ug/L	9035	USGS O-3116-87	80.4	75.0		30.5 - 119 *75 passed out of 77 reported results.*	Satisfactory
2,4-Dimethylphenol EPA Code: N/A	ug/L	9035	USGS O-3116-87	99.7	97.0		27.6 - 166 *76 passed out of 77 reported results.*	Satisfactory
2,4-Dinitrophenol EPA Code: N/A	ug/L	9035	USGS O-3116-87	< 3.30	< 100 *		*75 passed out of 76 reported results.*	Satisfactory
2-Chlorophenol EPA Code: N/A	ug/L	9035	USGS O-3116-87	42.1	39.9		16.0 - 63.8 *75 passed out of 77 reported results.*	Satisfactory
2-Nitrophenol EPA Code: N/A	ug/L	9035	USGS O-3116-87	< 1.4	< 100 *		*74 passed out of 76 reported results.*	Satisfactory
4-Chloro-3-methylphenol EPA Code: N/A	ug/L	9035	USGS O-3116-87	139	142		66.7 - 217 *75 passed out of 77 reported results.*	Satisfactory
Phenol EPA Code: N/A	ug/L	9035	USGS O-3116-87	113	83.0		14.8 - 198 *74 passed out of 77 reported results.*	Satisfactory
Sample: Non Potable Water Base/Neutrals								
1,2,4-Trichlorobenzene EPA Code: N/A	ug/L	9036	USGS O-3116-87	101	92.2		29.5 - 155 *72 passed out of 75 reported results.*	Satisfactory