WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Proficiency Test Report

Lab Id: 11642
EPA Lab Code: CO00009

USGS NATIONAL WATER QUALITY LABORATORY
POB 25046-MS 407-BLDG 95
DENVER, CO 80225-0046
(303) 236-2000
Director: MR. MERLE SHOCKEY

Shipment Date : 24-Jul-2006
Closing Date : 31-Aug-2006
Score Date : 20-Sep-2006

This report may contain data that are not covered by the NVLAP accreditation.
** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
NVLAP Lab Code 200387-0. ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 295 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/ Target *	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water Chlorinated Hydrocarbon Pesticides								
4,4-DDE ** EPA Code: N/A	ug/L	9517	USGS O-1126-95	6.32	8.96 *	5.26 - 10.3	4.00 - 11.5 *70 passed out of 72 reported results.*	Satisfactory
alpha-BHC EPA Code: N/A	ug/L	9517	USGS O-1126-95	8.03	6.76		3.27 - 10.2 *70 passed out of 72 reported results.*	Satisfactory
Sample: Non Potable Water Phenols								
Phenols ** EPA Code: N/A	mg/L	9509	SM 18 5530 (USGS Only)	2.7869	2.28	1.60 - 2.95	1.26 - 3.29 *65 passed out of 69 reported results.*	Satisfactory
Sample: Non Potable Water Purgeables								
1,1,1-Trichloroethane ** EPA Code: N/A	ug/L	9513	USGS O-4127-96	24.4	22.9 *	17.2 - 28.1	14.5 - 30.8 *102 passed out of 105 reported results.*	Satisfactory
1,1,2,2-Tetrachloroethane EPA Code: N/A	ug/L	9513	USGS O-4127-96	77.4	75.5		42.9 - 108 *98 passed out of 103 reported results.*	Satisfactory
1,1,2-Trichloroethane EPA Code: N/A	ug/L	9513	USGS O-4127-96	< 0.4	< 25.0 *		*100 passed out of 103 reported results.*	Satisfactory
1,1-Dichloroethene EPA Code: N/A	ug/L	9513	USGS O-4127-96	63.0	67.8		35.6 - 100 *102 passed out of 104 reported results.*	Satisfactory
1,2-Dichloroethane ** EPA Code: N/A	ug/L	9513	USGS O-4127-96	41.5	37.4 *	29.9 - 45.8	25.9 - 49.7 *97 passed out of 102 reported results.*	Satisfactory
1,2-Dichlorobenzene ** EPA Code: N/A	ug/L	9513	USGS O-4127-96	85.0	82.8 *	65.9 - 99.0	57.6 - 107 *120 passed out of 123 reported results.*	Satisfactory
1,2-Dichloropropane EPA Code: N/A	ug/L	9513	USGS O-4127-96	28.6	28.0		18.1 - 37.9 *101 passed out of 103 reported results.*	Satisfactory
1,3-Dichlorobenzene ** EPA Code: N/A	ug/L	9513	USGS O-4127-96	67.4	65.0 *	50.5 - 76.3	44.0 - 82.8 *121 passed out of 122 reported results.*	Satisfactory