

**WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM**

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER LAB
 EPA Lab Code: POB 25046-MS 407-BLDG 95
 CO00009 DENVER, CO-80225-0046
 (303) 236-3500
 Director: MR. GREGORY MOHRMAN

Shipment Date : 22-Jul-2002
 Closing Date : 05-Sep-2002
 Score Date : 24-Sep-2002

This report may contain data that are not covered by the NVLAP accreditation.
 ** indicates NVLAP accredited analyte. Lab Code 200387-0. ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 255 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/Target	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water		Organophosphate Pesticides						
Azinphos methyl EPA Code: N/A	ug/L	5512	Method Not Specified	8.39	7.88		D.L. - 16.3 11 passed out of 11 reported results.	Satisfactory
Diazinon EPA Code: N/A	ug/L	5512	Method Not Specified	9.05	7.83		3.66 - 12 11 passed out of 11 reported results.	Satisfactory
Disulfoton EPA Code: N/A	ug/L	5512	Method Not Specified	10.6	9.73		6.31 - 13.2 10 passed out of 11 reported results.	Satisfactory
Malathion EPA Code: N/A	ug/L	5512	Method Not Specified	7.88	7.5		1.64 - 13.4 11 passed out of 13 reported results.	Satisfactory
Parathion ethyl EPA Code: N/A	ug/L	5512	Method Not Specified	15.5	14.7		5.49 - 23.9 13 passed out of 14 reported results.	Satisfactory
Parathion methyl EPA Code: N/A	ug/L	5512	Method Not Specified	9.86	13.4		5.99 - 20.8 12 passed out of 13 reported results.	Satisfactory
Sample: Non Potable Water		Chlorinated Hydrocarbon Pesticides						
4,4-DDE ** EPA Code: N/A	ug/L	5517	Method Not Specified	< 0.05	< 5			Satisfactory 85 passed out of 98 reported results.